



MHRA Drug & Device Updates



Drug Safety Update

Nasal decongestant sprays and drops containing xylometazoline hydrochloride / oxymetazoline hydrochloride: increased risk of rebound congestion, rhinitis medicamentosa, and tachyphylaxis with overuse



Drug Safety Update

Finasteride and Dutasteride – updated safety warnings for psychiatric side effects and sexual dysfunction



Device Safety Information

Kimal Procedure Packs containing recalled components: Namic Angiographic Syringe with the risk of syringe disconnection; Namic Manifolds with the risk of foreign particulates. Important guidance for use in urgent procedures where there are no alternatives



Device Safety Information

Risk of severe harm from use of incorrect giving (administration) set for blood transfusion



Device Safety Information

Allurion Gastric Balloon: Updated safety information due to the risks of gastric outlet obstruction, small bowel obstruction and gastric perforation